Where can I find the results of the proficiency test?


The results of the proficiency test evaluation can be found in the user profile under the booked proficiency tests.

Report and certificate

RVEP 251035 Enterococcus milk powder

Documents

File	
Report RVEP 251035 23.04.2025	Save
Certificate (de) RVEP 251035 23.04.2025	Save
Certificate (en) RVEP 251035 23.04.2025	Save
Accompanying letter (de) RVEP 251035 10.03.2025	Save
Accompanying letter (en) RVEP 251035 10.03.2025	Save
Announcement (de) RVEP 251035 24.02.2025	Save
Announcement (en) RVEP 251035 24.02.2025	Save

The documents are available for download under View results. Your results are recorded in the report with a laboratory code. You will find your laboratory code on page 1 of the certificate.

results
aerobic total count sample 1 and sample 2

assigned value x_{pt} [log 10 cfu/g]					3,528								
± expanded uncertainty (k=t) of the assigned value $U(x_{pt})$					0,070								

Deutsches Referenzbüro für Ringversuche und Referenzmaterialien

DRRR-Proficiency Testing
RVEP 251035 Enterococcus milk powder

	customer data						performance					
	sample 1		sample 2			lab specific measurement uncertainty	sample 1			sample 2		
lab code no.	single value 1	single value 2	single value 1	single value 2	method description / standard		lab mean	lab mean	z'-score	lab mean	lab mean	z'-score
	[cfu/g]					[log 10 cfu/g]	[cfu/g]	[log 10 cfu/g]		[cfu/g]	[log 10 cfu/g]	
1	3360	4140	2040	2490	PCA, Plattenguss ISO 4833-1	0,063	3750	3,572	0,31	2265	3,353	-1,22
2	n/a	n/a	n/a	n/a								
3	2800	2400	3800	3300	PCA, Spatelverfahren ISO 4833.2	0,045	2600	3,414	-0,80	3550	3,549	0,15

Dear participant

Thank you very much for participating in our proficiency testing.

Your lab code number is: **3**